Certificate of Conformity

VWR CENTRIFUGE TUBE, CONICAL BOTTOM 15ml, HIGH-PERFORMANCE

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionés respectent toutes les specifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **525-0400**

Lot No: **239CB**

Material: **POLYPROPYLENE made of 100% virgin resin**

Sterility assurance level: **between$10^{-5}$and $10^{-6,}$ Sterile**

Properties: **Free of endotoxin to a detection level of 0.05EU/ml (Pyrogen-free)**
Free of detectable Nuclease contaminants as confirmed by a gel assay.
Free of Human DNA to a detectionlevel of 1 picogram DNA.
Heavy metals contamination of used resins, less then 1ppm (acc. USP (231))

This document has been produced electronically and is valid without a signature

Date of issue: **January 2013**

VWR International bvba/sprl, Researchpark Haasrode 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium
http://www.vwr.com

ISO 9001
ISO 14001
OHSAS 18001
BUREAU VERITAS
Certification