



Certificate of Conformity

VWR FILTER TIP, INDIVIDUALLY WRAPPED, ULTRAFINE™ STERILE

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **732-0538**

Lot No: **246C4**

Sterility assurance level: **between 10^{-5} and 10^{-6}**

Material: **POLYPROPYLENE made of 100% virgin resin**

Tested for heavy metals contamination using the U.S.P. (231) method and found to have results lower than 1ppm.

Properties: **Free of detectable Nuclease contaminants as confirmed by gel assay.**

Free of Human DNA to a detection level of 1 pico-gram of DNA.

Tested and found to be free of Endotoxin to a detection level of 0.05EU/mL

The sterilization program is an ISO 11137 validated process, with bioburden studies and radiation validation audits performed quarterly.

This document has been produced electronically and is valid without a signature

Date of issue: **April 2013**