



Certificate of Conformity

VWR CRUCIBLES

VWR International certifies that the products specified conform to all specifications and quality criteria mentioned in this document.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité mentionnés dans ce document.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle in diesem Dokument erwähnten Spezifikationen und Qualitätskriterien erfüllen.

Catalog No.	Type	Capacity (ml)	Ø×H (mm)	Pk
459-0194	Low form	5	30×19	5
459-0195	Low form	10	35×22	5
459-0196	Low form	17	40×25	5
459-0197	Low form	21	45×28	5
459-0198	Low form	34	50×32	5
459-0199	Low form	62	60×38	5
459-0200	Low form	91	70×44	5
459-0201	Medium form	10	30×25	5
459-0202	Medium form	12	35×28	5
459-0203	Medium form	20	40×32	5
459-0204	Medium form	30	45×36	5
459-0205	Medium form	45	50×40	5
459-0206	Medium form	80	60×48	5
459-0207	Medium form	120	70×56	5
459-0208	Medium form	200	80×62	5
459-0209	Tall form	15	30×38	5
459-0210	Tall form	26	35×44	5
459-0211	Tall form	35	40×50	5
459-0212	Tall form	50	45×56	5
459-0213	Tall form	72	50×62	5
459-0214	Tall form	130	60×75	5

Intended uses: Laboratory applications

Material: Porcelain C110

Shape: Low, middle or high form, without spout and graduation



Certificate of Conformity

Compliant to: ISO material for the manufacture lies within the group C110 according to CSN EN 60672 –3
Product dimension tolerances are controlled by DIN 40680
Acid resistance acc. to ISO 1775

Properties:

Unglazed products can be used at temperatures of up to 1350°C / 2462°F and glazed ones at temperatures of up to 1000°C / 1,832°F
Products have high resistance to acids and alkali
Density 2.4 g/cm³
Bending strength of unglazed products (70MPa)
Thermal conductivity (W.m⁻¹.K⁻¹) = 1.3

The compliance of demands is proved by regular tests and followed noticed tests from an independent laboratory:

Test according ČSN EN 60672-2
Determination of thermal shock resistance
Test in liquid nitrogen
Test resistance to acids and bases

Disclaimer: VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.

Date of issue: **April 2015**